

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062922\
 Data File : VU049508.D
 Acq On : 29 Jun 2022 13:26
 Operator : SY/MD
 Sample : N3482-10
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TB062322

Quant Time: Jun 30 00:48:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 28 10:35:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	274861	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.250	114	478230	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	432747	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.813	152	194272	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	186955	44.166	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	88.340%
35) Dibromofluoromethane	5.289	113	151808	45.372	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.740%
50) Toluene-d8	7.896	98	577610	48.402	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.800%
62) 4-Bromofluorobenzene	10.632	95	203096	44.428	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.860%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062922\
Data File : VU049508.D
Acq On : 29 Jun 2022 13:26
Operator : SY/MD
Sample : N3482-10
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TB062322

Quant Time: Jun 30 00:48:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 28 10:35:43 2022
Response via : Initial Calibration

