

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062922\
 Data File : VU049511.D
 Acq On : 29 Jun 2022 14:37
 Operator : SY/MD
 Sample : N3457-03DL 5X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW179D2-20220617DL

Quant Time: Jun 30 00:48:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 28 10:35:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	271349	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.250	114	477022	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	427697	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	190952	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	187588	44.889	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	89.780%
35) Dibromofluoromethane	5.288	113	150747	45.169	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.340%
50) Toluene-d8	7.896	98	573979	48.219	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.440%
62) 4-Bromofluorobenzene	10.632	95	203048	44.530	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.060%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.578	96	1309	0.444	ug/l	# 78
27) cis-1,2-Dichloroethene	4.668	96	21466	5.848	ug/l	99
44) Trichloroethene	6.542	130	161493	46.498	ug/l	89
64) Tetrachloroethene	8.552	164	12065	4.398	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062922\
Data File : VU049511.D
Acq On : 29 Jun 2022 14:37
Operator : SY/MD
Sample : N3457-03DL 5X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MW179D2-20220617DL

Quant Time: Jun 30 00:48:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 28 10:35:43 2022
Response via : Initial Calibration

