

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062922\
 Data File : VU049512.D
 Acq On : 29 Jun 2022 15:01
 Operator : SY/MD
 Sample : N3457-06DL 5X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 RE115D1-20220620DL

Quant Time: Jun 30 00:48:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 28 10:35:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.375	168	272291	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.250	114	480334	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	435274	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	193941	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	187700	44.760	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	89.520%		
35) Dibromofluoromethane	5.288	113	151442	45.065	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	90.120%		
50) Toluene-d8	7.899	98	579321	48.333	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	96.660%		
62) 4-Bromofluorobenzene	10.632	95	204912	44.629	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	89.260%		
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.581	101	7784	2.577	ug/l	92
12) 1,1-Dichloroethene	2.581	96	3228	1.091	ug/l	98
27) cis-1,2-Dichloroethene	4.671	96	1402	0.381	ug/l	71
30) Chloroform	5.092	83	1704	0.263	ug/l	92
38) Carbon Tetrachloride	5.530	117	1693	0.331	ug/l #	73
44) Trichloroethene	6.542	130	144083	41.199	ug/l	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062922\
Data File : VU049512.D
Acq On : 29 Jun 2022 15:01
Operator : SY/MD
Sample : N3457-06DL 5X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
RE115D1-20220620DL

Quant Time: Jun 30 00:48:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 28 10:35:43 2022
Response via : Initial Calibration

