

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062922\
 Data File : VU049516.D
 Acq On : 29 Jun 2022 16:35
 Operator : SY/MD
 Sample : N3457-16DL 5X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 RE125D2-20220621DL

Quant Time: Jun 30 00:49:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 28 10:35:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	264979	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.250	114	465051	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	418251	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	190956	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	184030	45.096	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.200%
35) Dibromofluoromethane	5.289	113	147280	45.267	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.540%
50) Toluene-d8	7.896	98	554815	47.809	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.620%
62) 4-Bromofluorobenzene	10.632	95	197712	44.476	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.960%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.578	101	7318	2.490	ug/l	93
12) 1,1-Dichloroethene	2.578	96	1997	0.694	ug/l	98
27) cis-1,2-Dichloroethene	4.674	96	2406	0.671	ug/l	73
44) Trichloroethene	6.542	130	177275	52.356	ug/l	84
64) Tetrachloroethene	8.555	164	3317	1.236	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062922\
Data File : VU049516.D
Acq On : 29 Jun 2022 16:35
Operator : SY/MD
Sample : N3457-16DL 5X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
RE125D2-20220621DL

Quant Time: Jun 30 00:49:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 28 10:35:43 2022
Response via : Initial Calibration

