

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062922\
 Data File : VU049538.D
 Acq On : 30 Jun 2022 02:21
 Operator : SY/MD
 Sample : N3522-01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BP-TT-TB-20220624

Quant Time: Jun 30 04:13:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 28 10:35:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	244015	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.250	114	428436	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	382716	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	169011	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	171866	45.733	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.460%
35) Dibromofluoromethane	5.288	113	135482	45.199	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.400%
50) Toluene-d8	7.896	98	515868	48.252	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.500%
62) 4-Bromofluorobenzene	10.632	95	176723	43.151	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.300%
Target Compounds						
30) Chloroform	5.089	83	8788	1.512	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062922\
Data File : VU049538.D
Acq On : 30 Jun 2022 02:21
Operator : SY/MD
Sample : N3522-01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BP-TT-TB-20220624

Quant Time: Jun 30 04:13:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 28 10:35:43 2022
Response via : Initial Calibration

