

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062922\
 Data File : VU049541.D
 Acq On : 30 Jun 2022 03:32
 Operator : SY/MD
 Sample : N3517-01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW-5-20220627

Quant Time: Jun 30 04:14:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 28 10:35:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	241461	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.247	114	426100	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	382270	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	163297	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	173604	46.685	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.360%
35) Dibromofluoromethane	5.289	113	138101	46.325	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.660%
50) Toluene-d8	7.896	98	512526	48.202	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.400%
62) 4-Bromofluorobenzene	10.632	95	177210	43.508	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.020%
Target Compounds						
					Qvalue	
30) Chloroform	5.086	83	1619	0.282	ug/l	92
64) Tetrachloroethene	8.555	164	3426	1.397	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062922\
Data File : VU049541.D
Acq On : 30 Jun 2022 03:32
Operator : SY/MD
Sample : N3517-01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MW-5-20220627

Quant Time: Jun 30 04:14:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 28 10:35:43 2022
Response via : Initial Calibration

