

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062922\
 Data File : VU049549.D
 Acq On : 30 Jun 2022 06:39
 Operator : SY/MD
 Sample : N3476-02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 52422

Quant Time: Jun 30 09:14:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 28 10:35:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	241159	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.250	114	431399	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	401684	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	201204	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	171519	46.182	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.360%
35) Dibromofluoromethane	5.289	113	135146	44.777	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.560%
50) Toluene-d8	7.896	98	527241	48.977	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.960%
62) 4-Bromofluorobenzene	10.632	95	201378	48.834	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.660%
Target Compounds						
16) Acetone	2.636	43	10497	5.605	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062922\
Data File : VU049549.D
Acq On : 30 Jun 2022 06:39
Operator : SY/MD
Sample : N3476-02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
52422

Quant Time: Jun 30 09:14:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 28 10:35:43 2022
Response via : Initial Calibration

