


Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062922\
Data File : VU049528.D
Acq On : 29 Jun 2022 21:02
Operator : SY/MD
Sample : BFB
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
Title : SW846 8260
Last Update : Tue Jun 28 10:35:43 2022



AutoFind: Scans 2889, 2890, 2891; Background Corrected with Scan 2876

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.4	11393	PASS
75	95	30	60	50.7	31341	PASS
95	95	100	100	100.0	61819	PASS
96	95	5	9	7.0	4328	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	74.7	46160	PASS
175	174	5	9	6.9	3202	PASS
176	174	95	101	98.7	45541	PASS
177	176	5	9	7.2	3287	PASS