

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062924\
 Data File : VU059620.D
 Acq On : 28 Jun 2024 15:07
 Operator : MD/SY
 Sample : P3006-22
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 JMRJ8

Quant Time: Jun 28 23:57:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061724WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jun 28 23:55:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	147537	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	153756	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	66845	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	39387	3.839	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	76.800%
7) Chloroethane-d5	1.907	69	42295	4.285	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	85.600%
11) 1,1-Dichloroethene-d2	2.560	65	18290	4.297	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	86.000%
20) 2-Butanone-d5	4.621	46	125508	59.392	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	118.780%
24) Chloroform-d	5.055	84	99043	4.703	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.000%
26) 1,2-Dichloroethane-d4	5.695	65	51157	5.043	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.800%
32) Benzene-d6	5.721	84	199304	4.710	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.200%
36) 1,2-Dichloropropane-d6	6.682	67	63522	4.899	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.000%
41) Toluene-d8	7.891	98	168840	4.292	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.800%
43) trans-1,3-Dichloroprop...	8.177	79	19487	4.799	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	96.000%
46) 2-Hexanone-d5	8.627	63	100731	59.708	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	119.420%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	57425	5.360	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	107.200%
66) 1,2-Dichlorobenzene-d4	12.187	152	64691	5.437	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	108.800%
Target Compounds						Qvalue
13) Acetone	2.634	43	5778	4.215	ug/L	95
47) Tetrachloroethene	8.547	164	68617	6.993	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062924\
Data File : VU059620.D
Acq On : 28 Jun 2024 15:07
Operator : MD/SY
Sample : P3006-22
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
JMRJ8

Quant Time: Jun 28 23:57:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061724WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jun 28 23:55:24 2024
Response via : Initial Calibration

