

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062924\
 Data File : VU059633.D
 Acq On : 28 Jun 2024 20:17
 Operator : MD/SY
 Sample : P3068-03
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GCL21

Quant Time: Jun 29 00:00:28 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061724WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jun 28 23:55:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	141074	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	152049	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	65983	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	36883	3.760	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	75.200%
7) Chloroethane-d5	1.908	69	40397	4.280	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	85.600%
11) 1,1-Dichloroethene-d2	2.560	65	16341	4.015	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	80.400%
20) 2-Butanone-d5	4.618	46	112122	55.488	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	110.980%
24) Chloroform-d	5.055	84	93760	4.657	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.200%
26) 1,2-Dichloroethane-d4	5.695	65	48661	5.017	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.400%
32) Benzene-d6	5.721	84	185245	4.427	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.600%
36) 1,2-Dichloropropane-d6	6.686	67	59655	4.652	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.000%
41) Toluene-d8	7.891	98	151640	3.898	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.000%
43) trans-1,3-Dichloroprop...	8.177	79	17682	4.404	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	88.000%
46) 2-Hexanone-d5	8.628	63	89347	53.555	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	107.100%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	54718	5.165	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	103.200%
66) 1,2-Dichlorobenzene-d4	12.190	152	59074	5.030	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.600%
Target Compounds						
					Qvalue	
9) Trichlorofluoromethane	2.136	101	6286	0.367	ug/L	94
34) Trichloroethene	6.538	95	27885	2.230	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062924\
Data File : VU059633.D
Acq On : 28 Jun 2024 20:17
Operator : MD/SY
Sample : P3068-03
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GCL21

Quant Time: Jun 29 00:00:28 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061724WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jun 28 23:55:24 2024
Response via : Initial Calibration

