




Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062924\
Data File : VU059650.D
Acq On : 29 Jun 2024 03:00
Operator : MD/SY
Sample : P3059-16
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
E1GP3

Manual IntegrationsAPPROVED

Quant Time: Jun 29 04:14:15 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061724WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jun 29 01:51:23 2024
Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 07/01/2024
Supervised By :Semsettin Yesilyurt 07/01/2024



TIC: VU059650.D\data.ms

(13) Acetone (T)

2.782min (+ 0.148) 1610.38 ug/L

response 2144666

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	34.00	4.23
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062924\
Data File : VU059650.D
Acq On : 29 Jun 2024 03:00
Operator : MD/SY
Sample : P3059-16
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
E1GP3

Manual IntegrationsAPPROVED

Quant Time: Jun 29 04:14:15 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061724WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jun 29 01:51:23 2024
Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 07/01/2024
Supervised By :Semsettin Yesilyurt 07/01/2024



TIC: VU059650.D\data.ms

(13) Acetone (T)

2.631min (-0.003) 10.13 ug/L m

response 13490

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	34.00	672.18#
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062924\
 Data File : VU059650.D
 Acq On : 29 Jun 2024 03:00
 Operator : MD/SY
 Sample : P3059-16
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 E1GP3

Manual IntegrationsAPPROVED

Quant Time: Jun 29 04:14:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061724WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Jun 29 01:51:23 2024
 Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 07/01/2024
 Supervised By :Semsettin Yesilyurt 07/01/2024

Compound	R.T.	Q	Ion	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene	6.242	114		143336	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117		148389	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152		65413	5.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3	1.596	65		33491	3.360	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	67.200%	
7) Chloroethane-d5	1.907	69		37567	3.917	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	78.400%	
11) 1,1-Dichloroethene-d2	2.560	65		15376	3.719	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	74.400%	
20) 2-Butanone-d5	4.618	46		100866	49.130	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	98.260%	
24) Chloroform-d	5.052	84		89904	4.395	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.800%	
26) 1,2-Dichloroethane-d4	5.695	65		44814	4.548	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.000%	
32) Benzene-d6	5.721	84		181468	4.443	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.800%	
36) 1,2-Dichloropropane-d6	6.685	67		57178	4.569	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	91.400%	
41) Toluene-d8	7.894	98		147787	3.893	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	77.800%	
43) trans-1,3-Dichloroprop...	8.174	79		15979	4.078	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.600%	
46) 2-Hexanone-d5	8.627	63		76726	47.124	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	94.240%	
56) 1,1,2,2-Tetrachloroeth...	10.749	84		48982	4.737	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.800%	
66) 1,2-Dichlorobenzene-d4	12.187	152		59940	5.148	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.000%	
Target Compounds							
13) Acetone	2.631	43		13490m	10.129	ug/L	
16) Methylene chloride	3.036	84		3120	0.259	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU062924\
Data File : VU059650.D
Acq On : 29 Jun 2024 03:00
Operator : MD/SY
Sample : P3059-16
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
E1GP3

Manual IntegrationsAPPROVED

Reviewed By :Mahesh Dadoda 07/01/2024
Supervised By :Semsettin Yesilyurt 07/01/2024

Quant Time: Jun 29 04:14:15 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061724WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jun 29 01:51:23 2024
Response via : Initial Calibration

