

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU063018\
 Data File : VU025080.D
 Acq On : 30 Jun 2018 11:55
 Operator : MD/SY
 Sample : VU0630WBL01
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0630WBL01

Quant Time: Jul 02 06:30:42 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	239207	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	352600	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	327573	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	171713	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	176279	45.14	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.28%	
35) Dibromofluoromethane	4.89	113	137714	47.06	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.12%	
50) Toluene-d8	7.57	98	502484	47.08	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.16%	
62) 4-Bromofluorobenzene	10.31	95	189835	44.79	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.58%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU063018\
Data File : VU025080.D
Acq On : 30 Jun 2018 11:55
Operator : MD/SY
Sample : VU0630WBL01
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0630WBL01

Quant Time: Jul 02 06:30:42 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

