

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU063018\
 Data File : VU025091.D
 Acq On : 30 Jun 2018 16:22
 Operator : MD/SY
 Sample : J3759-02
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 WP021MW177-180625

Quant Time: Jul 02 09:40:56 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	210464	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	319339	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	292306	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	150521	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	160639	46.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.50%	
35) Dibromofluoromethane	4.89	113	123519	46.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.22%	
50) Toluene-d8	7.57	98	453989	46.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.92%	
62) 4-Bromofluorobenzene	10.31	95	167608	43.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.34%	
Target Compounds						
4) Vinyl Chloride	1.40	62	2068	0.75	ug/l	98
27) cis-1,2-Dichloroethene	4.24	96	4270	1.41	ug/l	88
65) Chlorobenzene	9.12	112	20617	2.76	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU063018\
Data File : VU025091.D
Acq On : 30 Jun 2018 16:22
Operator : MD/SY
Sample : J3759-02
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WP021MW177-180625

Quant Time: Jul 02 09:40:56 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

