

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU063018\
 Data File : VU025095.D
 Acq On : 30 Jun 2018 18:00
 Operator : MD/SY
 Sample : J3759-06
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 WP021MW349-180626

Quant Time: Jul 02 09:49:58 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	212218	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	315824	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	291937	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	151136	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.31	65	162361	46.86	ug/l	0.00
Spiked Amount			Recovery	=	93.72%	
35) Dibromofluoromethane	4.89	113	124388	47.46	ug/l	0.00
Spiked Amount			Recovery	=	94.92%	
50) Toluene-d8	7.57	98	451164	47.19	ug/l	0.00
Spiked Amount			Recovery	=	94.38%	
62) 4-Bromofluorobenzene	10.31	95	170482	44.91	ug/l	0.00
Spiked Amount			Recovery	=	89.82%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Vinyl Chloride	1.40	62	22907	8.18	ug/l	94
12) 1,1-Dichloroethene	2.29	96	1781	0.74	ug/l	95
21) trans-1,2-Dichloroethene	2.99	96	1349	0.51	ug/l	92
27) cis-1,2-Dichloroethene	4.23	96	254825	83.28	ug/l	88
65) Chlorobenzene	9.12	112	15228	2.04	ug/l	93
91) 1,2-Dichlorobenzene	11.89	146	2572	0.47	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU063018\
Data File : VU025095.D
Acq On : 30 Jun 2018 18:00
Operator : MD/SY
Sample : J3759-06
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WP021MW349-180626

Quant Time: Jul 02 09:49:58 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

