

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU063018\
 Data File : VU025098.D
 Acq On : 30 Jun 2018 19:13
 Operator : MD/SY
 Sample : J3759-10
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DUP-0474-180626

Quant Time: Jul 02 07:17:41 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	207511	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	308411	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	289396	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	150339	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.31	65	158605	46.82	ug/l	0.00
Spiked Amount			Recovery	=	93.64%	
35) Dibromofluoromethane	4.89	113	122011	47.67	ug/l	0.00
Spiked Amount			Recovery	=	95.34%	
50) Toluene-d8	7.57	98	448659	48.06	ug/l	0.00
Spiked Amount			Recovery	=	96.12%	
62) 4-Bromofluorobenzene	10.31	95	168488	45.45	ug/l	0.00
Spiked Amount			Recovery	=	90.90%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Vinyl Chloride	1.40	62	21739	7.94	ug/l	95
12) 1,1-Dichloroethene	2.29	96	1656	0.70	ug/l	85
21) trans-1,2-Dichloroethene	2.99	96	1338	0.51	ug/l	81
27) cis-1,2-Dichloroethene	4.23	96	245583	82.08	ug/l	89
65) Chlorobenzene	9.12	112	13725	1.86	ug/l	96
91) 1,2-Dichlorobenzene	11.89	146	2524	0.47	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU063018\
Data File : VU025098.D
Acq On : 30 Jun 2018 19:13
Operator : MD/SY
Sample : J3759-10
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DUP-0474-180626

Quant Time: Jul 02 07:17:41 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

