

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU063018\
 Data File : VU025113.D
 Acq On : 01 Jul 2018 02:30
 Operator : MD/SY
 Sample : J3789-10
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TRIP-BLANK

Quant Time: Jul 02 09:21:26 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	215936	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	319532	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	295109	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	151282	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	161934	45.94	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.88%	
35) Dibromofluoromethane	4.89	113	124812	47.07	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.14%	
50) Toluene-d8	7.57	98	462517	47.82	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.64%	
62) 4-Bromofluorobenzene	10.31	95	167889	43.71	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.42%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU063018\
Data File : VU025113.D
Acq On : 01 Jul 2018 02:30
Operator : MD/SY
Sample : J3789-10
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TRIP-BLANK

Quant Time: Jul 02 09:21:26 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

