

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU063018\
 Data File : VU025121.D
 Acq On : 01 Jul 2018 05:44
 Operator : MD/SY
 Sample : J3804-07
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GPW-01

Quant Time: Jul 02 08:18:39 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	205886	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	305231	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	280531	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	148551	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	157367	46.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.64%	
35) Dibromofluoromethane	4.89	113	120458	47.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.10%	
50) Toluene-d8	7.57	98	439603	47.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.16%	
62) 4-Bromofluorobenzene	10.31	95	161713	44.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.16%	
Target Compounds						
16) Acetone	2.32	43	18025	9.98	ug/l	97
27) cis-1,2-Dichloroethene	4.23	96	1352	0.46	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU063018\
Data File : VU025121.D
Acq On : 01 Jul 2018 05:44
Operator : MD/SY
Sample : J3804-07
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GPW-01

Quant Time: Jul 02 08:18:39 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

