

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU063018\
 Data File : VU025125.D
 Acq On : 01 Jul 2018 07:21
 Operator : MD/SY
 Sample : J3789-03
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW-3

Quant Time: Jul 02 08:28:48 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	206121	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	307665	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	286558	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	147807	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	156299	46.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.90%	
35) Dibromofluoromethane	4.89	113	119711	46.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.76%	
50) Toluene-d8	7.57	98	443376	47.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.20%	
62) 4-Bromofluorobenzene	10.31	95	163756	44.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.56%	
Target Compounds						
16) Acetone	2.32	43	4312	2.38	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU063018\
Data File : VU025125.D
Acq On : 01 Jul 2018 07:21
Operator : MD/SY
Sample : J3789-03
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MW-3

Quant Time: Jul 02 08:28:48 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

