

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU063018\
 Data File : VU025129.D
 Acq On : 01 Jul 2018 08:58
 Operator : MD/SY
 Sample : J3789-07
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 55 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW-7

Quant Time: Jul 02 08:35:57 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	213987	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	315870	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	293892	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	149760	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.31	65	162915	46.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.28%	
35) Dibromofluoromethane	4.89	113	124249	47.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.80%	
50) Toluene-d8	7.57	98	456547	47.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.50%	
62) 4-Bromofluorobenzene	10.31	95	168490	44.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.76%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
11) Tert butyl alcohol	2.82	59	770731	907.15	ug/l	100
16) Acetone	2.32	43	4423	2.36	ug/l	99
19) Methyl tert-butyl Ether	3.00	73	3271801	357.22	ug/l	96
22) Diisopropyl ether	3.57	45	48403	5.40	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU063018\
Data File : VU025129.D
Acq On : 01 Jul 2018 08:58
Operator : MD/SY
Sample : J3789-07
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 55 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MW-7

Quant Time: Jul 02 08:35:57 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

