

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU063020\
 Data File : VU039249.D
 Acq On : 30 Jun 2020 23:13
 Operator : SY/MD
 Sample : VU0630WBL02
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK26

Quant Time: Jul 01 02:37:37 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR063020WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Jul 01 02:28:19 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	214207	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	227689	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	98340	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	68519	4.63	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	92.60%
7) Chloroethane-d5	1.93	69	82070	6.17	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	123.40%
11) 1,1-Dichloroethene-d2	2.59	63	92544	3.19	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	63.80%
20) 2-Butanone-d5	4.67	46	292756	48.89	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	97.78%
24) Chloroform-d	5.10	84	146352	4.69	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.80%
26) 1,2-Dichloroethane-d4	5.74	65	93993	4.83	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.60%
32) Benzene-d6	5.76	84	257367	4.59	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.80%
36) 1,2-Dichloropropane-d6	6.72	67	87019	4.45	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	89.00%
41) Toluene-d8	7.92	98	192216	4.19	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.80%
43) trans-1,3-Dichloropropene-	8.20	79	37723	4.28	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.60%
46) 2-Hexanone-d5	8.66	63	206778	45.70	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	91.40%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	82434	4.30	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	86.00%
64) 1,2-Dichlorobenzene-d4	12.20	152	86877	4.87	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.40%

Target Compounds					Ovalue
16) Methylene chloride	3.08	84	5774	0.278 ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU063020\
Data File : VU039249.D
Acq On : 30 Jun 2020 23:13
Operator : SY/MD
Sample : VU0630WBL02
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK26

Quant Time: Jul 01 02:37:37 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR063020WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Jul 01 02:28:19 2020
Response via : Initial Calibration

