

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU063020\
 Data File : VU039269.D
 Acq On : 01 Jul 2020 08:52
 Operator : SY/MD
 Sample : L3134-09
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 ETYK6

Quant Time: Jul 01 10:46:21 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR063020WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Jul 01 08:49:01 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	183382	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.43	117	187921	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.82	152	84306	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	69438	5.49	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	109.80%
7) Chloroethane-d5	1.93	69	62128	5.45	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	109.00%
11) 1,1-Dichloroethene-d2	2.59	63	90950	3.66	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	73.20%
20) 2-Butanone-d5	4.67	46	161822	31.57	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	63.14%
24) Chloroform-d	5.10	84	116545	4.37	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.40%
26) 1,2-Dichloroethane-d4	5.74	65	71109	4.27	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.40%
32) Benzene-d6	5.76	84	211254	4.56	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.20%
36) 1,2-Dichloropropane-d6	6.72	67	65834	4.08	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	81.60%
41) Toluene-d8	7.92	98	176068	4.65	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.00%
43) trans-1,3-Dichloropropene-	8.20	79	32036	4.41	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	88.20%
46) 2-Hexanone-d5	8.65	63	106962	28.64	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	57.28%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	52319	3.31	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	66.20%
64) 1,2-Dichlorobenzene-d4	12.20	152	67989	4.45	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	89.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU063020\
Data File : VU039269.D
Acq On : 01 Jul 2020 08:52
Operator : SY/MD
Sample : L3134-09
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
ETVK6

Quant Time: Jul 01 10:46:21 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR063020WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Jul 01 08:49:01 2020
Response via : Initial Calibration

