

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU063022\
 Data File : VU049564.D
 Acq On : 30 Jun 2022 15:32
 Operator : SY/MD
 Sample : N3458-01DL 10X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW17811-20220621DL

Quant Time: Jul 01 05:25:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 28 10:35:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	243182	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.247	114	432812	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	393538	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	181511	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	172749	46.126	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.260%
35) Dibromofluoromethane	5.289	113	138296	45.671	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.340%
50) Toluene-d8	7.896	98	517254	47.893	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.780%
62) 4-Bromofluorobenzene	10.632	95	187007	45.201	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.400%
Target Compounds						
						Qvalue
27) cis-1,2-Dichloroethene	4.665	96	11580	3.520	ug/l	97
44) Trichloroethene	6.542	130	102821	32.629	ug/l	85
45) 1,2-Dichloropropane	6.790	63	8019	2.183	ug/l	98
64) Tetrachloroethene	8.555	164	4647	1.841	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU063022\
Data File : VU049564.D
Acq On : 30 Jun 2022 15:32
Operator : SY/MD
Sample : N3458-01DL 10X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MW17811-20220621DL

Quant Time: Jul 01 05:25:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 28 10:35:43 2022
Response via : Initial Calibration

