

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU063022\
 Data File : VU049571.D
 Acq On : 30 Jun 2022 18:19
 Operator : SY/MD
 Sample : N3554-14
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 152140-SPC-03-60

Quant Time: Jul 01 05:27:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 28 10:35:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	234277	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.250	114	412817	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	375425	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	173696	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	164206	45.511	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.020%
35) Dibromofluoromethane	5.289	113	129666	44.896	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.800%
50) Toluene-d8	7.896	98	496526	48.200	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.400%
62) 4-Bromofluorobenzene	10.632	95	179741	45.549	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.100%
Target Compounds						
44) Trichloroethene	6.543	130	1507	0.501	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU063022\
Data File : VU049571.D
Acq On : 30 Jun 2022 18:19
Operator : SY/MD
Sample : N3554-14
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
152140-SPC-03-60

Quant Time: Jul 01 05:27:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 28 10:35:43 2022
Response via : Initial Calibration

