

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU063022\
 Data File : VU049575.D
 Acq On : 30 Jun 2022 19:54
 Operator : SY/MD
 Sample : N3458-02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 RE122D1-20220621

Quant Time: Jul 01 05:27:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 28 10:35:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	231203	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.247	114	404354	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	362775	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	163473	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	164745	46.268	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.540%
35) Dibromofluoromethane	5.289	113	129339	45.720	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.440%
50) Toluene-d8	7.896	98	488025	48.366	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.740%
62) 4-Bromofluorobenzene	10.632	95	170334	44.069	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.140%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.578	101	3104	1.210	ug/l	100
12) 1,1-Dichloroethene	2.578	96	1147	0.457	ug/l	89
21) trans-1,2-Dichloroethene	3.353	96	857	0.317	ug/l	91
27) cis-1,2-Dichloroethene	4.668	96	5094	1.629	ug/l	98
44) Trichloroethene	6.542	130	818312	277.955	ug/l	87
64) Tetrachloroethene	8.552	164	3688	1.585	ug/l	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU063022\
Data File : VU049575.D
Acq On : 30 Jun 2022 19:54
Operator : SY/MD
Sample : N3458-02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
RE122D1-20220621

Quant Time: Jul 01 05:27:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 28 10:35:43 2022
Response via : Initial Calibration

