

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU063022\
 Data File : VU049576.D
 Acq On : 30 Jun 2022 20:17
 Operator : SY/MD
 Sample : N3458-10
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 RE120D2-20220622

Quant Time: Jul 01 05:27:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 28 10:35:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	228723	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.250	114	403433	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	364333	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	166167	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	163312	46.363	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.720%
35) Dibromofluoromethane	5.288	113	129279	45.803	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.600%
50) Toluene-d8	7.896	98	485791	48.255	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.500%
62) 4-Bromofluorobenzene	10.632	95	173374	44.957	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.920%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.578	101	34281	13.511	ug/l	93
12) 1,1-Dichloroethene	2.578	96	4865	1.958	ug/l	# 77
27) cis-1,2-Dichloroethene	4.668	96	7338	2.372	ug/l	97
38) Carbon Tetrachloride	5.523	117	1913	0.446	ug/l	# 89
44) Trichloroethene	6.542	130	1301890	443.221	ug/l	86
55) 1,1,2-Trichloroethane	8.401	97	959	0.316	ug/l	# 85
64) Tetrachloroethene	8.552	164	8965	3.836	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU063022\
Data File : VU049576.D
Acq On : 30 Jun 2022 20:17
Operator : SY/MD
Sample : N3458-10
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
RE120D2-20220622

Quant Time: Jul 01 05:27:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 28 10:35:43 2022
Response via : Initial Calibration

