

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070121\
 Data File : VU044290.D
 Acq On : 01 Jul 2021 16:51
 Operator : SY/MD
 Sample : M2905-23
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DBK46

Quant Time: Jul 02 01:53:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR062121WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jul 02 01:46:49 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.256	114	136135	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.423	117	126307	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.819	152	67621	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	29129	3.952	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	79.000%
7) Chloroethane-d5	1.919	69	32300	4.451	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.000%
11) 1,1-Dichloroethene-d2	2.571	65	15687	4.147	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	83.000%
20) 2-Butanone-d5	4.645	46	137086	50.507	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	101.020%
24) Chloroform-d	5.073	84	79091	4.767	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.400%
26) 1,2-Dichloroethane-d4	5.710	65	46930	4.732	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.600%
32) Benzene-d6	5.735	84	154375	4.921	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.400%
36) 1,2-Dichloropropane-d6	6.697	67	51513	5.067	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.400%
41) Toluene-d8	7.906	98	143900	4.944	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.800%
43) trans-1,3-Dichloroprop...	8.185	79	20077	4.258	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.200%
46) 2-Hexanone-d5	8.642	63	114772	51.101	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	102.200%
56) 1,1,2,2-Tetrachloroeth...	10.764	84	43288	4.790	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.800%
66) 1,2-Dichlorobenzene-d4	12.198	152	58291	5.232	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070121\
Data File : VU044290.D
Acq On : 01 Jul 2021 16:51
Operator : SY/MD
Sample : M2905-23
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DBK46

Quant Time: Jul 02 01:53:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR062121WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jul 02 01:46:49 2021
Response via : Initial Calibration

