

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070122\
 Data File : VU049592.D
 Acq On : 01 Jul 2022 19:18
 Operator : SY/MD
 Sample : N3569-18
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TB-06302022

Quant Time: Jul 04 06:49:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 28 10:35:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	199592	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.250	114	345450	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	320866	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	143555	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	145217	47.243	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.480%
35) Dibromofluoromethane	5.288	113	113144	46.815	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.620%
50) Toluene-d8	7.899	98	419215	48.631	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.260%
62) 4-Bromofluorobenzene	10.632	95	148335	44.921	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.840%
Target Compounds						
					Qvalue	
16) Acetone	2.629	43	2371	1.530	ug/l	93
64) Tetrachloroethene	8.555	164	10907	5.300	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070122\
Data File : VU049592.D
Acq On : 01 Jul 2022 19:18
Operator : SY/MD
Sample : N3569-18
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TB-06302022

Quant Time: Jul 04 06:49:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 28 10:35:43 2022
Response via : Initial Calibration

