

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070122\
 Data File : VU049595.D
 Acq On : 01 Jul 2022 20:31
 Operator : SY/MD
 Sample : N3564-14 10X
 Misc : 5.93g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 WC-14B-11-3-GR

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 07/05/2022
 Supervised By :Mahesh Dadoda 07/05/2022

Quant Time: Jul 04 06:50:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 28 10:35:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	204073	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.247	114	354469	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	317269	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	152732	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	143623	45.698	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	91.400%		
35) Dibromofluoromethane	5.289	113	113345	45.704	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	91.400%		
50) Toluene-d8	7.896	98	443149	50.100	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	100.200%		
62) 4-Bromofluorobenzene	10.632	95	182439	53.843	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	107.680%		
Target Compounds						
						Qvalue
31) Cyclohexane	5.385	56	261680	55.452	ug/l	85
39) Methylcyclohexane	6.761	83	620334	143.923	ug/l	89
67) Ethyl Benzene	9.571	91	36085	3.030	ug/l	99
68) m/p-Xylenes	9.690	106	37242	8.453	ug/l	83
73) Isopropylbenzene	10.485	105	110461	10.153	ug/l	97
78) n-propylbenzene	10.906	91	207567	15.700	ug/l	94
80) 1,3,5-Trimethylbenzene	11.089	105	157754	16.608	ug/l	94
83) tert-Butylbenzene	11.420	119	9396m	1.006	ug/l	
84) 1,2,4-Trimethylbenzene	11.468	105	51315	5.499	ug/l	95
85) sec-Butylbenzene	11.645	105	92323	7.860	ug/l	96
86) p-Isopropyltoluene	11.793	119	57949	6.101	ug/l	89
89) n-Butylbenzene	12.208	91	139631	15.880	ug/l #	58

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070122\
Data File : VU049595.D
Acq On : 01 Jul 2022 20:31
Operator : SY/MD
Sample : N3564-14 10X
Misc : 5.93g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WC-14B-11-3-GR

Quant Time: Jul 04 06:50:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 28 10:35:43 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 07/05/2022
Supervised By :Mahesh Dadoda 07/05/2022

