

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070122\
 Data File : VU049597.D
 Acq On : 01 Jul 2022 21:18
 Operator : SY/MD
 Sample : N3564-24 10X
 Misc : 4.17g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 WC-16-2-5-GR

Quant Time: Jul 04 06:51:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 28 10:35:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	244333	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.247	114	430598	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	398048	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	197494	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	170289	45.255	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.500%
35) Dibromofluoromethane	5.285	113	136333	45.255	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.500%
50) Toluene-d8	7.896	98	522385	48.616	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.240%
62) 4-Bromofluorobenzene	10.632	95	196851	47.825	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.640%
Target Compounds						
					Qvalue	
25) 2-Butanone	4.713	43	3495	1.229	ug/l	# 86
68) m/p-Xylenes	9.693	106	9549	1.728	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070122\
Data File : VU049597.D
Acq On : 01 Jul 2022 21:18
Operator : SY/MD
Sample : N3564-24 10X
Misc : 4.17g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WC-16-2-5-GR

Quant Time: Jul 04 06:51:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 28 10:35:43 2022
Response via : Initial Calibration

