

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070122\
 Data File : VU049596.D
 Acq On : 01 Jul 2022 20:54
 Operator : SY/MD
 Sample : N3564-19 10X
 Misc : 2.83g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 WC-16-2-6-GR

Quant Time: Jul 04 06:51:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 28 10:35:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.369	168	236419	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.247	114	418336	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	389526	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	185321	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	166152	45.634	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.260%
35) Dibromofluoromethane	5.289	113	132757	45.359	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.720%
50) Toluene-d8	7.896	98	512727	49.116	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.240%
62) 4-Bromofluorobenzene	10.632	95	190388	47.611	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.220%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070122\
Data File : VU049596.D
Acq On : 01 Jul 2022 20:54
Operator : SY/MD
Sample : N3564-19 10X
Misc : 2.83g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WC-16-2-6-GR

Quant Time: Jul 04 06:51:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 28 10:35:43 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.369 min Scan# 1253
Delta R.T. -0.000 min
Lab File: VU049596.D
Acq: 01 Jul 2022 20:54




Instrument :
MSVOA_U
ClientSampleId :
WC-16-2-6-GR


Tgt Ion:168 Resp: 236419
Ion Ratio Lower Upper
168 100
99 60.3 39.3 58.9#



#33
1,2-Dichloroethane-d4
Concen: 45.634 ug/l
RT: 5.700 min Scan# 1356
Delta R.T. 0.000 min
Lab File: VU049596.D
Acq: 01 Jul 2022 20:54

Tgt Ion: 65 Resp: 166152
Ion Ratio Lower Upper
65 100
67 54.2 0.0 108.0









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.247 min Scan# 114
Delta R.T. 0.000 min
Lab File: VU049596.D
Acq: 01 Jul 2022 20:54




Instrument :
MSVOA_U
ClientSampleId :
WC-16-2-6-GR


Tgt Ion:114 Resp: 418336
Ion Ratio Lower Upper
114 100
63 21.4 0.0 37.8
88 16.4 0.0 29.2



#35
Dibromofluoromethane
Concen: 45.359 ug/l
RT: 5.289 min Scan# 1228
Delta R.T. 0.000 min
Lab File: VU049596.D
Acq: 01 Jul 2022 20:54

Tgt Ion:113 Resp: 132757
Ion Ratio Lower Upper
113 100
111 103.7 82.1 123.1
192 18.6 18.0 27.0









#50
Toluene-d8
Concen: 49.116 ug/l
RT: 7.896 min Scan# 2039
Delta R.T. 0.000 min
Lab File: VU049596.D
Acq: 01 Jul 2022 20:54




Instrument :
MSVOA_U
ClientSampleId :
WC-16-2-6-GR


Tgt Ion: 98 Resp: 512727
Ion Ratio Lower Upper
98 100
100 64.8 51.7 77.5



#62
4-Bromofluorobenzene
Concen: 47.611 ug/l
RT: 10.632 min Scan# 2890
Delta R.T. -0.003 min
Lab File: VU049596.D
Acq: 01 Jul 2022 20:54

Tgt Ion: 95 Resp: 190388
Ion Ratio Lower Upper
95 100
174 75.2 0.0 182.0
176 72.1 0.0 177.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 9.417 min Scan# 2512
Delta R.T. -0.003 min
Lab File: VU049596.D
Acq: 01 Jul 2022 20:54

Instrument :
MSVOA_U
ClientSampleId :
WC-16-2-6-GR

Tgt Ion:117 Resp: 389526
Ion Ratio Lower Upper
117 100
82 59.4 41.7 62.5
119 31.6 25.7 38.5



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 11.812 min Scan# 3257
Delta R.T. 0.000 min
Lab File: VU049596.D
Acq: 01 Jul 2022 20:54

Tgt Ion:152 Resp: 185321
Ion Ratio Lower Upper
152 100
115 62.6 37.0 110.9
150 156.5 0.0 342.2

