

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070124\
 Data File : VU059673.D
 Acq On : 01 Jul 2024 10:35
 Operator : MD/SY
 Sample : P3068-15DL 5X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GCL31DL

Quant Time: Jul 01 11:04:33 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061724WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Jun 29 01:51:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	144238	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	152345	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	63325	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	28943	2.886	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	57.800%
7) Chloroethane-d5	1.907	69	33832	3.506	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	70.200%
11) 1,1-Dichloroethene-d2	2.560	65	12879	3.095	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	62.000%
20) 2-Butanone-d5	4.621	46	94823	45.898	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	91.800%
24) Chloroform-d	5.055	84	83332	4.048	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.000%
26) 1,2-Dichloroethane-d4	5.695	65	41168	4.151	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.000%
32) Benzene-d6	5.721	84	156022	3.721	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	74.400%
36) 1,2-Dichloropropane-d6	6.682	67	52756	4.106	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	82.200%
41) Toluene-d8	7.891	98	127345	3.267	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	65.400%#
43) trans-1,3-Dichloroprop...	8.177	79	15302	3.804	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	76.000%
46) 2-Hexanone-d5	8.627	63	71691	42.888	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	85.780%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	46893	4.418	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	88.400%
66) 1,2-Dichlorobenzene-d4	12.187	152	50270	4.460	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	89.200%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.383	85	6402	0.524	ug/L	98
9) Trichlorofluoromethane	2.132	101	114524	6.547	ug/L	100
10) 1,1,2-Trichloro-1,2,2-...	2.570	101	2793	0.252	ug/L #	86
16) Methylene chloride	3.039	84	4000	0.330	ug/L	83
34) Trichloroethene	6.537	95	4353	0.347	ug/L	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070124\
Data File : VU059673.D
Acq On : 01 Jul 2024 10:35
Operator : MD/SY
Sample : P3068-15DL 5X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GCL31DL

Quant Time: Jul 01 11:04:33 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061724WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jun 29 01:51:23 2024
Response via : Initial Calibration

