

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070124\
 Data File : VU059683.D
 Acq On : 01 Jul 2024 14:34
 Operator : MD/SY
 Sample : P3059-22
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Jul 02 01:39:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061724WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Jun 29 01:51:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	130763	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	137733	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	57979	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	30365	3.340	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	66.800%
7) Chloroethane-d5	1.908	69	35289	4.034	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	80.600%
11) 1,1-Dichloroethene-d2	2.560	65	14101	3.738	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	74.800%
20) 2-Butanone-d5	4.621	46	101507	54.196	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	108.400%
24) Chloroform-d	5.055	84	86022	4.609	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.200%
26) 1,2-Dichloroethane-d4	5.695	65	42422	4.719	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.400%
32) Benzene-d6	5.721	84	164533	4.341	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.800%
36) 1,2-Dichloropropane-d6	6.685	67	54076	4.656	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.200%
41) Toluene-d8	7.894	98	129631	3.679	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	73.600%
43) trans-1,3-Dichloroprop...	8.177	79	15635	4.299	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	86.000%
46) 2-Hexanone-d5	8.628	63	73441	48.596	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	97.200%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	48683	5.073	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	101.400%
66) 1,2-Dichlorobenzene-d4	12.187	152	52158	5.054	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.000%
Target Compounds						Qvalue
16) Methylene chloride	3.043	84	2029	0.185	ug/L	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070124\
Data File : VU059683.D
Acq On : 01 Jul 2024 14:34
Operator : MD/SY
Sample : P3059-22
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Jul 02 01:39:05 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061724WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jun 29 01:51:23 2024
Response via : Initial Calibration

