

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070124\
 Data File : VU059688.D
 Acq On : 01 Jul 2024 16:29
 Operator : MD/SY
 Sample : P2979-14
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_U
ClientSampleId :
 C0B41

Quant Time: Jul 02 01:39:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061724WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Jun 29 01:51:23 2024
 Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|-------------------------------|--------|----------------|----------|--------|----------|----------|
| ----- | | | | | | |
| Internal Standards | | | | | | |
| 1) 1,4-Difluorobenzene | 6.242 | 114 | 134357 | 5.000 | ug/L | 0.00 |
| 28) Chlorobenzene-d5 | 9.412 | 117 | 144197 | 5.000 | ug/L | 0.00 |
| 58) 1,4-Dichlorobenzene-d4 | 11.807 | 152 | 60695 | 5.000 | ug/L | 0.00 |
| System Monitoring Compounds | | | | | | |
| 4) Vinyl Chloride-d3 | 1.595 | 65 | 28718 | 3.074 | ug/L | 0.00 |
| Spiked Amount | 5.000 | Range 40 - 130 | Recovery | = | 61.400% | |
| 7) Chloroethane-d5 | 1.907 | 69 | 33208 | 3.694 | ug/L | 0.00 |
| Spiked Amount | 5.000 | Range 65 - 130 | Recovery | = | 73.800% | |
| 11) 1,1-Dichloroethene-d2 | 2.560 | 65 | 13485 | 3.479 | ug/L | 0.00 |
| Spiked Amount | 5.000 | Range 60 - 125 | Recovery | = | 69.600% | |
| 20) 2-Butanone-d5 | 4.618 | 46 | 103681 | 53.876 | ug/L | 0.00 |
| Spiked Amount | 50.000 | Range 40 - 130 | Recovery | = | 107.760% | |
| 24) Chloroform-d | 5.055 | 84 | 80044 | 4.174 | ug/L | 0.00 |
| Spiked Amount | 5.000 | Range 70 - 125 | Recovery | = | 83.400% | |
| 26) 1,2-Dichloroethane-d4 | 5.695 | 65 | 41360 | 4.478 | ug/L | 0.00 |
| Spiked Amount | 5.000 | Range 70 - 130 | Recovery | = | 89.600% | |
| 32) Benzene-d6 | 5.721 | 84 | 152608 | 3.845 | ug/L | 0.00 |
| Spiked Amount | 5.000 | Range 70 - 125 | Recovery | = | 77.000% | |
| 36) 1,2-Dichloropropane-d6 | 6.682 | 67 | 51418 | 4.228 | ug/L | 0.00 |
| Spiked Amount | 5.000 | Range 60 - 140 | Recovery | = | 84.600% | |
| 41) Toluene-d8 | 7.894 | 98 | 120918 | 3.278 | ug/L | 0.00 |
| Spiked Amount | 5.000 | Range 70 - 130 | Recovery | = | 65.600%# | |
| 43) trans-1,3-Dichloroprop... | 8.174 | 79 | 15273 | 4.011 | ug/L | 0.00 |
| Spiked Amount | 5.000 | Range 55 - 130 | Recovery | = | 80.200% | |
| 46) 2-Hexanone-d5 | 8.627 | 63 | 80212 | 50.697 | ug/L | 0.00 |
| Spiked Amount | 50.000 | Range 45 - 130 | Recovery | = | 101.400% | |
| 56) 1,1,2,2-Tetrachloroeth... | 10.749 | 84 | 49630 | 4.940 | ug/L | 0.00 |
| Spiked Amount | 5.000 | Range 65 - 120 | Recovery | = | 98.800% | |
| 66) 1,2-Dichlorobenzene-d4 | 12.187 | 152 | 48403 | 4.481 | ug/L | 0.00 |
| Spiked Amount | 5.000 | Range 80 - 120 | Recovery | = | 89.600% | |
| ----- | | | | | | |
| Target Compounds | | | | | | Qvalue |
| 16) Methylene chloride | 3.046 | 84 | 1696 | 0.150 | ug/L | 86 |
| ----- | | | | | | |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070124\
 Data File : VU059688.D
 Acq On : 01 Jul 2024 16:29
 Operator : MD/SY
 Sample : P2979-14
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0B41

Quant Time: Jul 02 01:39:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061724WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Jun 29 01:51:23 2024
 Response via : Initial Calibration

