

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070124\
Data File : VU059690.D
Acq On : 01 Jul 2024 17:17
Operator : MD/SY
Sample : P3084-03
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
E1GE2


Manual IntegrationsAPPROVED

Reviewed By :Mahesh Dadoda 07/05/2024
Supervised By :Semsettin Yesilyurt 07/05/2024

Quant Time: Jul 02 01:39:55 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061724WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jun 29 01:51:23 2024
Response via : Initial Calibration

07/05/2024
Supervised By :Semsettin
Yesilyurt

07/05/2024



Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070124\
Data File : VU059690.D
Acq On : 01 Jul 2024 17:17
Operator : MD/SY
Sample : P3084-03
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 22 Sample Multiplier: 1




Instrument :
MSVOA_U
ClientSampleId :
E1GE2

Manual IntegrationsAPPROVED

Reviewed By :Mahesh Dadoda 07/05/2024
Supervised By :Semsettin Yesilyurt 07/05/2024

Quant Time: Jul 02 01:39:55 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061724WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jun 29 01:51:23 2024
Response via : Initial Calibration

07/05/2024
Supervised By :Semsettin
Yesilyurt



TIC: VU059690.D\data.ms

(13) Acetone (T)

2.448min (-0.186) 157.91 ug/L

response 190173

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	34.00	6.30
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070124\
Data File : VU059690.D
Acq On : 01 Jul 2024 17:17
Operator : MD/SY
Sample : P3084-03
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 22 Sample Multiplier: 1




Instrument :
MSVOA_U
ClientSampleId :
E1GE2

Manual IntegrationsAPPROVED

Reviewed By :Mahesh Dadoda 07/05/2024
Supervised By :Semsettin Yesilyurt 07/05/2024

Quant Time: Jul 02 01:39:55 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061724WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jun 29 01:51:23 2024
Response via : Initial Calibration

07/05/2024
Supervised By :Semsettin
Yesilyurt



TIC: VU059690.D\data.ms

(13) Acetone (T)

2.631min (-0.003) 26.79 ug/L m

response 32265

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	34.00	37.12
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070124\
 Data File : VU059690.D
 Acq On : 01 Jul 2024 17:17
 Operator : MD/SY
 Sample : P3084-03
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 E1GE2

Manual IntegrationsAPPROVED

Reviewed By :Mahesh Dadoda 07/05/2024
 Supervised By :Semsettin Yesilyurt 07/05/2024

Quant Time: Jul 02 01:39:55 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061724WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Jun 29 01:51:23 2024
 Response via : Initial Calibration

07/05/2024
 Supervised By :Semsettin
 Yesilyurt

07/05/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	129619	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	142922	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.808	152	62088	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	30736	3.410	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	68.200%	
7) Chloroethane-d5	1.908	69	35456	4.088	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	81.800%	
11) 1,1-Dichloroethene-d2	2.561	65	14233	3.806	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	76.200%	
20) 2-Butanone-d5	4.618	46	115024	61.955	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	123.920%	
24) Chloroform-d	5.056	84	87816	4.747	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	95.000%	
26) 1,2-Dichloroethane-d4	5.695	65	45018	5.052	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	101.000%	
32) Benzene-d6	5.721	84	164205	4.175	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	83.400%	
36) 1,2-Dichloropropane-d6	6.683	67	55533	4.607	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	92.200%	
41) Toluene-d8	7.891	98	135008	3.692	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	73.800%	
43) trans-1,3-Dichloroprop...	8.174	79	16721	4.430	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	88.600%	
46) 2-Hexanone-d5	8.628	63	85558	54.558	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	109.120%	
56) 1,1,2,2-Tetrachloroeth...	10.750	84	52770	5.299	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	106.000%	
66) 1,2-Dichlorobenzene-d4	12.190	152	55823	5.052	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	101.000%	
Target Compounds						
13) Acetone	2.631	43	32265m	26.791	ug/L	
16) Methylene chloride	3.040	84	2771	0.254	ug/L	86
25) Chloroform	5.081	83	4159	0.224	ug/L #	72
42) Toluene	7.965	91	4904	0.107	ug/L	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed