

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070124\
Data File : VU059693.D
Acq On : 01 Jul 2024 18:29
Operator : MD/SY
Sample : P3084-06
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 25 Sample Multiplier: 1


Instrument :
MSVOA_U
ClientSampleId :
E1GC7

Manual IntegrationsAPPROVED

Reviewed By :Mahesh Dadoda 07/05/2024
Supervised By :Semsettin Yesilyurt 07/05/2024

Supervised By :Semsettin
Yesilyurt
07/05/2024

Quant Time: Jul 02 01:40:16 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061724WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jun 29 01:51:23 2024
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070124\
Data File : VU059693.D
Acq On : 01 Jul 2024 18:29
Operator : MD/SY
Sample : P3084-06
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
E1GC7




Manual IntegrationsAPPROVED

Reviewed By :Mahesh Dadoda 07/05/2024
Supervised By :Semsettin Yesilyurt 07/05/2024

Supervised By :Semsettin
Yesilyurt

07/05/2024

Quant Time: Jul 02 01:40:16 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061724WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jun 29 01:51:23 2024
Response via : Initial Calibration



TIC: VU059693.D\data.ms

(13) Acetone (T)

2.448min (-0.186) 106.81 ug/L

response 130191

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	34.00	9.36
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070124\
Data File : VU059693.D
Acq On : 01 Jul 2024 18:29
Operator : MD/SY
Sample : P3084-06
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 25 Sample Multiplier: 1




Instrument :
MSVOA_U
ClientSampleId :
E1GC7

Manual IntegrationsAPPROVED

Reviewed By :Mahesh Dadoda 07/05/2024
Supervised By :Semsettin Yesilyurt 07/05/2024

Supervised By :Semsettin
Yesilyurt

Quant Time: Jul 02 01:40:16 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061724WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jun 29 01:51:23 2024
Response via : Initial Calibration



TIC: VU059693.D\data.ms

(13) Acetone (T)

2.634min (+ 0.000) 26.76 ug/L m

response 32611

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	34.00	37.36
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070124\
 Data File : VU059693.D
 Acq On : 01 Jul 2024 18:29
 Operator : MD/SY
 Sample : P3084-06
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 E1GC7

Manual IntegrationsAPPROVED

Reviewed By :Mahesh Dadoda 07/05/2024
 Supervised By :Semsettin Yesilyurt 07/05/2024

Supervised By :Semsettin
 Yesilyurt
 07/05/2024

Quant Time: Jul 02 01:40:16 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061724WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Jun 29 01:51:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	131184	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	140057	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	61932	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	28438	3.118	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	62.400%
7) Chloroethane-d5	1.907	69	33078	3.769	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	75.400%
11) 1,1-Dichloroethene-d2	2.560	65	12859	3.398	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	68.000%
20) 2-Butanone-d5	4.621	46	111544	59.364	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	118.720%
24) Chloroform-d	5.055	84	82768	4.421	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.400%
26) 1,2-Dichloroethane-d4	5.695	65	42051	4.662	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.200%
32) Benzene-d6	5.721	84	154460	4.007	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	80.200%
36) 1,2-Dichloropropane-d6	6.685	67	52902	4.479	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	89.600%
41) Toluene-d8	7.894	98	127655	3.562	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	71.200%
43) trans-1,3-Dichloroprop...	8.177	79	15239	4.120	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	82.400%
46) 2-Hexanone-d5	8.627	63	84743	55.144	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	110.280%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	51604	5.288	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	105.800%
66) 1,2-Dichlorobenzene-d4	12.187	152	55231	5.011	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.200%
Target Compounds						
13) Acetone	2.634	43	32611m	26.755	ug/L	
16) Methylene chloride	3.042	84	3068	0.278	ug/L	92
25) Chloroform	5.075	83	4754	0.253	ug/L	99
42) Toluene	7.968	91	5719	0.127	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed