

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU070218\
 Data File : VU025154.D
 Acq On : 02 Jul 2018 19:41
 Operator : MD/SY
 Sample : J3245-22
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 CL-01-062918-C

Quant Time: Jul 03 03:02:02 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	203512	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	313677	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	284941	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	151498	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	154548	46.52	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.04%	
35) Dibromofluoromethane	4.89	113	121038	46.49	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.98%	
50) Toluene-d8	7.57	98	442429	46.59	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.18%	
62) 4-Bromofluorobenzene	10.31	95	164248	43.57	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.14%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) Chloromethane	1.33	50	659	0.253	ug/l #	87
6) Chloroethane	1.69	64	61	Below Cal	#	43
8) Diethyl Ether	2.10	74	172	Below Cal	#	1
11) Tert butyl alcohol	2.84	59	1639	2.028	ug/l #	72
13) Acrolein	2.19	56	94	0.222	ug/l #	1
16) Acetone	2.32	43	17203	9.637	ug/l	100
18) Methyl Acetate	2.62	43	21360	5.582	ug/l	98
20) Methylene Chloride	2.70	84	254687	90.928	ug/l	98
25) 2-Butanone	4.28	43	2448	1.021	ug/l #	82
29) Tetrahydrofuran	4.66	42	725	0.488	ug/l #	74
31) Cyclohexane	4.99	56	4059	Below Cal	#	1
36) 1,1-Dichloropropene	4.99	75	15386	4.174	ug/l #	51
38) Carbon Tetrachloride	4.99	117	21026	5.187	ug/l #	16
43) Isopropyl Acetate	5.77	43	35385	5.096	ug/l #	62
44) Trichloroethene	6.19	130	19000	6.356	ug/l	95
51) 4-Methyl-2-Pentanone	7.46	43	1462	0.335	ug/l #	66
52) Toluene	7.64	92	2248	0.333	ug/l	96
60) Dibromochloromethane	8.23	129	2036	0.594	ug/l #	11
64) Tetrachloroethene	8.23	164	2307	0.852	ug/l	86
67) Ethyl Benzene	9.26	91	3670	0.288	ug/l	95
68) m/p-Xylenes	9.38	106	6623	1.352	ug/l	93
69) o-Xylene	9.79	106	2355	0.486	ug/l	99
76) 1,2,3-Trichloropropane	10.31	75	86655	25.980	ug/l #	35
81) trans-1,4-Dichloro-2-buten	10.31	75	86655	60.129	ug/l #	100
84) 1,2,4-Trimethylbenzene	11.15	105	4509	0.439	ug/l	93
85) sec-Butylbenzene	11.15	105	4509	0.371	ug/l #	57
95) Naphthalene	13.75	128	128769	12.082	ug/l	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU070218\
Data File : VU025154.D
Acq On : 02 Jul 2018 19:41
Operator : MD/SY
Sample : J3245-22
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
CL-01-062918-C

Quant Time: Jul 03 03:02:02 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 4.99 min Scan# 1368

Delta R.T. -0.00 min

Lab File: VU025154.D

Acq: 02 Jul 2018 19:41

Instrument :

MSVOA_U

ClientSampleId :


CL-01-062918-C

Tot Ion:168 Resp: 203512

Ion Ratio Lower Upper




168 100

99 56.5 43.2 64.8



Abundance Ion 167.90 (167.60 to 168.60): VU024502.D

Ion 98.90 (98.60 to 99.60): VU024502.D



#3

Chloromethane

Concen: 0.253 ug/l

RT: 1.33 min Scan# 230

Delta R.T. -0.00 min

Lab File: VU025154.D


Acq: 02 Jul 2018 19:41

Tgt Ion: 50 Resp: 659

Ion Ratio Lower Upper



50 100


52 25.6 26.4 39.6#



Abundance Ion 49.90 (49.60 to 50.60): VU024502.D

Ion 51.90 (51.60 to 52.60): VU024502.D





#6

Chloroethane

Concen: Below Cal

RT: 1.69 min Scan# 341

Delta R.T. -0.01 min

Lab File: VU025154.D

Acq: 02 Jul 2018 19:41

Instrument :

MSVOA_U

ClientSampleId :


CL-01-062918-C

Tot Ion: 64 Resp: 61

Ion Ratio Lower Upper

64 100

66 0.0 25.5 38.3#



Abundance Ion 63.90 (63.60 to 64.60): VU0

Ion 65.90 (65.60 to 66.60): VU0

1.69

120

100

80

60



40

20

0

1.68 1.69 1.70

Time-->



#8

Diethyl Ether

Concen: Below Cal

RT: 2.10 min Scan# 469

Delta R.T. -0.01 min

Lab File: VU025154.D


Acq: 02 Jul 2018 19:41

Tgt Ion: 74 Resp: 172

Ion Ratio Lower Upper

74 100

45 265.1 55.0 165.0#



Abundance Ion 74.00 (73.70 to 74.70): VU0

Ion 45.00 (44.70 to 45.70): VU0

500

400

300


200


100

0

2.08 2.09 2.10 2.11

Time-->









#11

Tert butyl alcohol
 Concen: 2.028 ug/l
 RT: 2.84 min Scan# 699
 Delta R.T. 0.01 min
 Lab File: VU025154.D
 Acq: 02 Jul 2018 19:41

Instrument :
 MSVOA_U
 ClientSampleId :
 CL-01-062918-C




Tot Ion: 59 Resp: 1639
 Ion Ratio Lower Upper
 59 100
 57 0.0 8.5 12.7#




#13

Acrolein
 Concen: 0.222 ug/l
 RT: 2.19 min Scan# 498
 Delta R.T. -0.00 min
 Lab File: VU025154.D
 Acq: 02 Jul 2018 19:41

Tgt Ion: 56 Resp: 94
 Ion Ratio Lower Upper
 56 100
 55 168.1 58.4 87.6#





#16

Acetone

Concen: 9.637 ug/l

RT: 2.32 min Scan# 539

Delta R.T. -0.00 min

Lab File: VU025154.D

Acq: 02 Jul 2018 19:41

Instrument :

MSVOA_U

ClientSampleId :


CL-01-062918-C

Tot Ion: 43 Resp: 17203

Ion Ratio Lower Upper

43 100

58 30.7 24.4 36.6



Abundance Ion 43.00 (42.70 to 43.70): VU0

Ion 58.00 (57.70 to 58.70): VU0

12000

10000

8000

6000



4000

2000

0

Time--> 2.25 2.30 2.35 2.40

2.32



#18

Methyl Acetate

Concen: 5.582 ug/l

RT: 2.62 min Scan# 631

Delta R.T. 0.00 min

Lab File: VU025154.D


Acq: 02 Jul 2018 19:41

Tgt Ion: 43 Resp: 21360

Ion Ratio Lower Upper

43 100

74 23.2 18.0 27.0



Abundance Ion 43.00 (42.70 to 43.70): VU0

Ion 74.00 (73.70 to 74.70): VU0


10000


5000

0

Time--> 2.55 2.60 2.65 2.70

2.62





#20

Methvlene Chloride

Concen: 90.928 ug/l

RT: 2.70 min Scan# 657

Delta R.T. -0.00 min

Lab File: VU025154.D

Acq: 02 Jul 2018 19:41

Instrument :

MSVOA_U

ClientSampleId :

CL-01-062918-C

Tot Ion: 84 Resp: 254687


Ion Ratio Lower Upper

84 100

49 125.9 103.8 155.8

51 38.5 32.0 48.0

86 64.4 51.6 77.4






Abundance Ion 83.90 (83.60 to 84.60): VU0

Ion 48.90 (48.60 to 49.60): VU0

Ion 50.90 (50.60 to 51.60): VU0

Ion 85.90 (85.60 to 86.60): VU0



#25

2-Butanone

Concen: 1.021 ug/l

RT: 4.28 min Scan# 1149

Delta R.T. 0.02 min

Lab File: VU025154.D


Acq: 02 Jul 2018 19:41

Tgt Ion: 43 Resp: 2448

Ion Ratio Lower Upper



43 100


72 15.4 19.6 29.4#



Abundance Ion 43.00 (42.70 to 43.70): VU0

Ion 72.00 (71.70 to 72.70): VU0





#29

Tetrahydrofuran

Concen: 0.488 ug/l

RT: 4.66 min Scan# 1265

Delta R.T. 0.02 min

Lab File: VU025154.D

Acq: 02 Jul 2018 19:41

Instrument :

MSVOA_U

ClientSampleId :

CL-01-062918-C





Tot Ion: 42 Resp: 725

Ion Ratio Lower Upper

42 100

72 21.7 34.5 51.7#

71 29.2 32.2 48.4#



#31

Cyclohexane

Concen: Below Cal

RT: 4.99 min Scan# 1367

Delta R.T. -0.02 min

Lab File: VU025154.D

Acq: 02 Jul 2018 19:41




Tgt Ion: 56 Resp: 4059


Ion Ratio Lower Upper

56 100

69 201.0 24.8 37.2#

84 128.3 65.2 97.8#









#33
 1,2-Dichloroethane-d4
 Concen: 46.517 ug/l
 RT: 5.31 min Scan# 1469
 Delta R.T. -0.00 min
 Lab File: VU025154.D
 Acq: 02 Jul 2018 19:41




Instrument :
 MSVOA_U
 ClientSampleId :
 CL-01-062918-C


Tot Ion: 65 Resp: 154548
 Ion Ratio Lower Upper
 65 100
 67 51.3 0.0 107.0



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 5.89 min Scan# 1649
 Delta R.T. -0.00 min
 Lab File: VU025154.D
 Acq: 02 Jul 2018 19:41

Tgt Ion: 114 Resp: 313677
 Ion Ratio Lower Upper
 114 100
 63 20.7 0.0 45.4
 88 16.1 0.0 31.0









#35
 Dibromofluoromethane
 Concen: 46.494 ug/l
 RT: 4.89 min Scan# 1336
 Delta R.T. -0.00 min
 Lab File: VU025154.D
 Acq: 02 Jul 2018 19:41




Instrument :
 MSVOA_U
 ClientSampleId :
 CL-01-062918-C


Tot Ion	Ratio	Lower	Upper
113	100		
111	103.1	82.2	123.4
192	20.0	16.2	24.4



#36
 1,1-Dichloropropene
 Concen: 4.174 ug/l
 RT: 4.99 min Scan# 1367
 Delta R.T. -0.16 min
 Lab File: VU025154.D
 Acq: 02 Jul 2018 19:41

Tgt Ion	Ratio	Lower	Upper
75	100		
110	9.2	17.0	50.9#
77	0.0	24.2	36.4#





#38

Carbon Tetrachloride

Concen: 5.187 ug/l

RT: 4.99 min Scan# 1368

Delta R.T. -0.15 min

Lab File: VU025154.D

Acq: 02 Jul 2018 19:41

Instrument :

MSVOA_U

ClientSampleId :

CL-01-062918-C


Tot Ion:117 Resp: 21026

Ion Ratio Lower Upper

117 100

119 5.0 76.1 114.1#




121 0.0 23.7 35.5#



Abundance Ion 116.80 (116.50 to 117.50): V

Ion 118.80 (118.50 to 119.50): V

Ion 120.80 (120.50 to 121.50): V



#43

Isopropyl Acetate

Concen: 5.096 ug/l

RT: 5.77 min Scan# 1611

Delta R.T. 0.22 min

Lab File: VU025154.D

Acq: 02 Jul 2018 19:41


Tgt Ion: 43 Resp: 35385

Ion Ratio Lower Upper

43 100

61 0.4 15.4 23.0#



87 0.0 10.0 15.0#




Abundance Ion 43.00 (42.70 to 43.70): VU0

Ion 61.00 (60.70 to 61.70): VU0

Ion 87.00 (86.70 to 87.70): VU0





#44

Trichloroethene

Concen: 6.356 ug/l

RT: 6.19 min Scan# 1742

Delta R.T. -0.00 min

Lab File: VU025154.D

Acq: 02 Jul 2018 19:41

Instrument :

MSVOA_U

ClientSampleId :


CL-01-062918-C

Tot Ion:130 Resp: 19000

Ion Ratio Lower Upper




130 100

95 95.7 0.0 202.4



Abundance Ion 129.80 (129.50 to 130.50): V

Ion 94.90 (94.60 to 95.60): VU0



#50

Toluene-d8

Concen: 46.593 ug/l

RT: 7.57 min Scan# 2171

Delta R.T. -0.00 min

Lab File: VU025154.D


Acq: 02 Jul 2018 19:41

Tgt Ion: 98 Resp: 442429

Ion Ratio Lower Upper



98 100


100 63.0 51.1 76.7



Abundance Ion 98.00 (97.70 to 98.70): VU0

Ion 100.00 (99.70 to 100.70): VU0









#51
 4-Methyl-2-Pentanone
 Concen: 0.335 ug/l
 RT: 7.46 min Scan# 2138
 Delta R.T. 0.00 min
 Lab File: VU025154.D
 Acq: 02 Jul 2018 19:41




Instrument :
 MSVOA_U
 ClientSampleId :
 CL-01-062918-C


Tot Ion: 43 Resp: 1462
 Ion Ratio Lower Upper
 43 100
 58 16.9 30.0 45.0#



#52
 Toluene
 Concen: 0.333 ug/l
 RT: 7.64 min Scan# 2193
 Delta R.T. -0.00 min
 Lab File: VU025154.D
 Acq: 02 Jul 2018 19:41

Tgt Ion: 92 Resp: 2248
 Ion Ratio Lower Upper
 92 100
 91 181.6 140.5 210.7





#60

Dibromochloromethane

Concen: 0.594 ug/l

RT: 8.23 min Scan# 2376

Delta R.T. -0.25 min

Lab File: VU025154.D

Acq: 02 Jul 2018 19:41

Instrument :

MSVOA_U

ClientSampleId :


CL-01-062918-C

Tot Ion:129 Resp: 2036

Ion Ratio Lower Upper




129 100

127 0.0 38.6 116.0#



Abundance Ion 128.80 (128.50 to 129.50): V

Ion 126.80 (126.50 to 127.50): V



#62

4-Bromofluorobenzene

Concen: 43.565 ug/l

RT: 10.31 min Scan# 3023

Delta R.T. -0.00 min

Lab File: VU025154.D

Acq: 02 Jul 2018 19:41


Tgt Ion: 95 Resp: 164248

Ion Ratio Lower Upper

95 100

174 84.3 0.0 165.8



176 81.8 0.0 159.4




Abundance Ion 94.90 (94.60 to 95.60): VU0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 9.09 min Scan# 2645
Delta R.T. -0.00 min
Lab File: VU025154.D
Acq: 02 Jul 2018 19:41




Instrument :
MSVOA_U
ClientSampleId :
CL-01-062918-C


Tot Ion:117 Resp: 284941
Ion Ratio Lower Upper
117 100
82 56.6 44.3 66.5
119 32.4 25.4 38.2



#64
Tetrachloroethene
Concen: 0.852 ug/l
RT: 8.23 min Scan# 2377
Delta R.T. 0.00 min
Lab File: VU025154.D
Acq: 02 Jul 2018 19:41

Tgt Ion:164 Resp: 2307
Ion Ratio Lower Upper
164 100
166 112.5 101.7 152.5
129 80.8 76.9 115.3
131 80.2 74.9 112.3









#67
Ethyl Benzene
Concen: 0.288 ug/l
RT: 9.26 min Scan# 2695
Delta R.T. 0.00 min
Lab File: VU025154.D
Acq: 02 Jul 2018 19:41




Instrument :
MSVOA_U
ClientSampleId :
CL-01-062918-C


Tot Ion: 91 Resp: 3670
Ion Ratio Lower Upper
91 100
106 33.2 24.2 36.4



#68
m/p-Xylenes
Concen: 1.352 ug/l
RT: 9.38 min Scan# 2733
Delta R.T. -0.00 min
Lab File: VU025154.D
Acq: 02 Jul 2018 19:41

Tgt Ion: 106 Resp: 6623
Ion Ratio Lower Upper
106 100
91 196.5 166.5 249.7









#69
o-Xylene
Concen: 0.486 ug/l
RT: 9.79 min Scan# 2861
Delta R.T. 0.01 min
Lab File: VU025154.D
Acq: 02 Jul 2018 19:41




Instrument :
MSVOA_U
ClientSampleId :
CL-01-062918-C


Tot Ion:106 Resp: 2355
Ion Ratio Lower Upper
106 100
91 223.6 110.7 331.9



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 11.49 min Scan# 3389
Delta R.T. -0.00 min
Lab File: VU025154.D
Acq: 02 Jul 2018 19:41

Tgt Ion:152 Resp: 151498
Ion Ratio Lower Upper
152 100
115 56.5 43.0 129.0
150 155.3 0.0 354.0









#76
 1,2,3-Trichloropropane
 Concen: 25.980 ug/l
 RT: 10.31 min Scan# 3023
 Delta R.T. -0.19 min
 Lab File: VU025154.D
 Acq: 02 Jul 2018 19:41




Instrument :
 MSVOA_U
 ClientSampleId :
 CL-01-062918-C


Tot Ion: 75 Resp: 86655
 Ion Ratio Lower Upper
 75 100
 77 1.1 20.9 62.7#



#81
 trans-1,4-Dichloro-2-butene
 Concen: 60.129 ug/l
 RT: 10.31 min Scan# 3023
 Delta R.T. -0.20 min
 Lab File: VU025154.D
 Acq: 02 Jul 2018 19:41

Tgt Ion: 75 Resp: 86655
 Ion Ratio Lower Upper
 75 100
 53 0.0 0.0 0.0
 89 0.0 0.0 0.0









#84
 1,2,4-Trimethylbenzene
 Concen: 0.439 ug/l
 RT: 11.15 min Scan# 3285
 Delta R.T. 0.00 min
 Lab File: VU025154.D
 Acq: 02 Jul 2018 19:41




Instrument :
 MSVOA_U
 ClientSampleId :
 CL-01-062918-C




Tot Ion:105 Resp: 4509
 Ion Ratio Lower Upper
 105 100
 120 40.7 22.6 67.8



#85
 sec-Butylbenzene
 Concen: 0.371 ug/l
 RT: 11.15 min Scan# 3285
 Delta R.T. -0.17 min
 Lab File: VU025154.D
 Acq: 02 Jul 2018 19:41

Tgt Ion:105 Resp: 4509
 Ion Ratio Lower Upper
 105 100
 134 0.0 9.7 29.1#





#95

Naphthalene

Concen: 12.082 ug/l

RT: 13.75 min Scan# 4092

Delta R.T. 0.00 min

Lab File: VU025154.D

Acq: 02 Jul 2018 19:41

Instrument :

MSVOA_U

ClientSampleId :

CL-01-062918-C

Tot Ion:128 Resp: 128769

Ion Ratio Lower Upper

128 100

127 12.7 10.6 16.0

129 11.0 8.6 12.8

