

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU070218\
 Data File : VU025143.D
 Acq On : 02 Jul 2018 15:11
 Operator : MD/SY
 Sample : J3759-17DL 5X
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DUP-0475-180626DL

Quant Time: Jul 03 04:59:49 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	206667	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	308068	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	284916	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	148267	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.31	65	156471	46.38	ug/l	0.00
Spiked Amount			Recovery	=	92.76%	
35) Dibromofluoromethane	4.89	113	121450	47.50	ug/l	0.00
Spiked Amount			Recovery	=	95.00%	
50) Toluene-d8	7.57	98	442849	47.49	ug/l	0.00
Spiked Amount			Recovery	=	94.98%	
62) 4-Bromofluorobenzene	10.31	95	163108	44.05	ug/l	0.00
Spiked Amount			Recovery	=	88.10%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Vinyl Chloride	1.40	62	129229	47.41	ug/l	97
27) cis-1,2-Dichloroethene	4.23	96	87285	29.29	ug/l	88
44) Trichloroethene	6.19	130	6932	2.36	ug/l	94
65) Chlorobenzene	9.12	112	128034	17.59	ug/l	98
83) tert-Butylbenzene	11.10	119	3006	0.32	ug/l	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU070218\
Data File : VU025143.D
Acq On : 02 Jul 2018 15:11
Operator : MD/SY
Sample : J3759-17DL 5X
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DUP-0475-180626DL

Quant Time: Jul 03 04:59:49 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

