

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU070218\
 Data File : VU025147.D
 Acq On : 02 Jul 2018 16:49
 Operator : MD/SY
 Sample : J3248-10
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EO-03-062818-B

Quant Time: Jul 03 05:14:22 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	206289	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	313796	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	291178	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	155691	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.31	65	159604	47.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.78%	
35) Dibromofluoromethane	4.89	113	124025	47.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.24%	
50) Toluene-d8	7.57	98	449059	47.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.54%	
62) 4-Bromofluorobenzene	10.31	95	166676	44.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.38%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
11) Tert butyl alcohol	2.83	59	5652	6.90	ug/l #	94
16) Acetone	2.32	43	40526	22.40	ug/l	97
18) Methyl Acetate	2.62	43	19296	4.97	ug/l	97
20) Methylene Chloride	2.71	84	924086	325.48	ug/l	98
25) 2-Butanone	4.27	43	7648	3.15	ug/l #	86
95) Naphthalene	13.75	128	146659	13.32	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU070218\
Data File : VU025147.D
Acq On : 02 Jul 2018 16:49
Operator : MD/SY
Sample : J3248-10
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EO-03-062818-B

Quant Time: Jul 03 05:14:22 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

