

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU070218\
 Data File : VU025153.D
 Acq On : 02 Jul 2018 19:16
 Operator : MD/SY
 Sample : J3245-18
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 CL-01-062918-A

Quant Time: Jul 03 05:41:48 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	202210	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	305255	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	283784	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	152006	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	153903	46.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.24%	
35) Dibromofluoromethane	4.89	113	118345	46.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.42%	
50) Toluene-d8	7.57	98	438897	47.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.00%	
62) 4-Bromofluorobenzene	10.31	95	165095	45.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.00%	
Target Compounds						
16) Acetone	2.32	43	15850	8.94	ug/l	95
18) Methyl Acetate	2.62	43	20951	5.51	ug/l	96
20) Methylene Chloride	2.70	84	63418	22.79	ug/l	97
68) m/p-Xylenes	9.38	106	9455	1.94	ug/l	98
84) 1,2,4-Trimethylbenzene	11.15	105	26982	2.62	ug/l	98
95) Naphthalene	13.75	128	38407	4.07	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU070218\
Data File : VU025153.D
Acq On : 02 Jul 2018 19:16
Operator : MD/SY
Sample : J3245-18
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
CL-01-062918-A

Quant Time: Jul 03 05:41:48 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

