

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU070218\  
 Data File : VU025161.D  
 Acq On : 02 Jul 2018 23:21  
 Operator : MD/SY  
 Sample : VU0702WBL02  
 Misc : 5.0mL/MSVOA U/WATER  
 ALS Vial : 32 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 ClientSampleId :  
 VU0702WBL02

Quant Time: Jul 03 06:48:29 2018  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_U\METHOD\82U061318W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Jun 13 13:55:26 2018  
 Response via : Initial Calibration

| Internal Standards         | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|----------------------------|-------|------|----------|-------|-------|----------|
| 1) Pentafluorobenzene      | 4.99  | 168  | 206189   | 50.00 | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene    | 5.89  | 114  | 309670   | 50.00 | ug/l  | 0.00     |
| 63) Chlorobenzene-d5       | 9.09  | 117  | 287904   | 50.00 | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4 | 11.49 | 152  | 146815   | 50.00 | ug/l  | 0.00     |

| System Monitoring Compounds | R.T.  | QIon | Response | Conc   | Units  | Dev(Min) |
|-----------------------------|-------|------|----------|--------|--------|----------|
| 33) 1,2-Dichloroethane-d4   | 5.31  | 65   | 155577   | 46.22  | ug/l   | 0.00     |
| Spiked Amount               |       |      |          | 50.000 |        |          |
|                             |       |      | Recovery | =      | 92.44% |          |
| 35) Dibromofluoromethane    | 4.89  | 113  | 120517   | 46.89  | ug/l   | 0.00     |
| Spiked Amount               |       |      |          | 50.000 |        |          |
|                             |       |      | Recovery | =      | 93.78% |          |
| 50) Toluene-d8              | 7.57  | 98   | 445906   | 47.57  | ug/l   | 0.00     |
| Spiked Amount               |       |      |          | 50.000 |        |          |
|                             |       |      | Recovery | =      | 95.14% |          |
| 62) 4-Bromofluorobenzene    | 10.31 | 95   | 163315   | 43.88  | ug/l   | 0.00     |
| Spiked Amount               |       |      |          | 50.000 |        |          |
|                             |       |      | Recovery | =      | 87.76% |          |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU070218\  
Data File : VU025161.D  
Acq On : 02 Jul 2018 23:21  
Operator : MD/SY  
Sample : VU0702WBL02  
Misc : 5.0mL/MSVOA U/WATER  
ALS Vial : 32 Sample Multiplier: 1

Instrument :  
MSVOA\_U  
ClientSampleId :  
VU0702WBL02

Quant Time: Jul 03 06:48:29 2018  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_U\METHOD\82U061318W.M  
Quant Title : SW846 8260  
QLast Update : Wed Jun 13 13:55:26 2018  
Response via : Initial Calibration

