

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU070219\
 Data File : VU033142.D
 Acq On : 02 Jul 2019 15:46
 Operator : JC/SP
 Sample : K3608-05
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GAML4

Quant Time: Jul 03 02:56:28 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR070219WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Jul 03 02:14:33 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	88508	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.08	117	91731	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	43204	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	36553	7.14	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	142.80%#
7) Chloroethane-d5	1.68	69	29274	6.25	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	125.00%
11) 1,1-Dichloroethene-d2	2.27	63	55575	4.60	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	92.00%
20) 2-Butanone-d5	4.19	46	100569	53.92	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	107.84%
24) Chloroform-d	4.64	84	67337	5.60	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	112.00%
26) 1,2-Dichloroethane-d4	5.31	65	37892	5.46	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	109.20%
32) Benzene-d6	5.33	84	125411	5.70	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	114.00%
36) 1,2-Dichloropropane-d6	6.32	67	40769	5.57	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	111.40%
41) Toluene-d8	7.56	98	108995	5.15	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.00%
43) trans-1,3-Dichloropropene-	7.85	79	15597	4.95	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	99.00%
46) 2-Hexanone-d5	8.31	63	62434	48.09	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	96.18%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	32876	5.11	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	102.20%
64) 1,2-Dichlorobenzene-d4	11.85	152	44898	5.77	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	115.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU070219\
Data File : VU033142.D
Acq On : 02 Jul 2019 15:46
Operator : JC/SP
Sample : K3608-05
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GAML4

Quant Time: Jul 03 02:56:28 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR070219WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Jul 03 02:14:33 2019
Response via : Initial Calibration

