

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU070219\
 Data File : VU033143.D
 Acq On : 02 Jul 2019 16:10
 Operator : JC/SP
 Sample : K3608-06
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GAML5

Quant Time: Jul 03 02:58:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR070219WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Jul 03 02:14:33 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	90248	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	92455	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	43203	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	36466	6.98	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	139.60%#
7) Chloroethane-d5	1.68	69	29292	6.13	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	122.60%
11) 1,1-Dichloroethene-d2	2.27	63	55537	4.51	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	90.20%
20) 2-Butanone-d5	4.19	46	99024	52.07	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.14%
24) Chloroform-d	4.64	84	66933	5.46	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	109.20%
26) 1,2-Dichloroethane-d4	5.31	65	38087	5.38	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.60%
32) Benzene-d6	5.33	84	124321	5.60	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	112.00%
36) 1,2-Dichloropropane-d6	6.32	67	40573	5.50	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	110.00%
41) Toluene-d8	7.56	98	109098	5.11	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.20%
43) trans-1,3-Dichloropropene-	7.85	79	15522	4.89	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	97.80%
46) 2-Hexanone-d5	8.31	63	61189	46.76	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.52%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	32462	5.01	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.20%
64) 1,2-Dichlorobenzene-d4	11.85	152	43366	5.58	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	111.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU070219\
Data File : VU033143.D
Acq On : 02 Jul 2019 16:10
Operator : JC/SP
Sample : K3608-06
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GAML5

Quant Time: Jul 03 02:58:37 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR070219WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Jul 03 02:14:33 2019
Response via : Initial Calibration

