

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU070219\
 Data File : VU033144.D
 Acq On : 02 Jul 2019 16:34
 Operator : JC/SP
 Sample : K3608-07
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GAML6

Quant Time: Jul 03 03:00:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR070219WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Jul 03 02:14:33 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	85102	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.08	117	88910	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	41640	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	36486	7.41	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	148.20%#
7) Chloroethane-d5	1.68	69	29463	6.54	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	130.80%#
11) 1,1-Dichloroethene-d2	2.27	63	55633	4.79	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	95.80%
20) 2-Butanone-d5	4.19	46	102785	57.31	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	114.62%
24) Chloroform-d	4.64	84	66418	5.75	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	115.00%
26) 1,2-Dichloroethane-d4	5.30	65	38446	5.76	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	115.20%
32) Benzene-d6	5.33	84	126117	5.91	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	118.20%
36) 1,2-Dichloropropane-d6	6.32	67	40521	5.71	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	114.20%
41) Toluene-d8	7.56	98	110988	5.41	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	108.20%
43) trans-1,3-Dichloropropene-	7.85	79	15423	5.05	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	101.00%
46) 2-Hexanone-d5	8.31	63	63467	50.43	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	100.86%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	32894	5.28	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	105.60%
64) 1,2-Dichlorobenzene-d4	11.85	152	43805	5.84	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	116.80%

Target Compounds

					Ovalue
13) Acetone	2.32	43	1947	1.350 ug/L	82
34) Trichloroethene	6.18	95	193888	27.126 ug/L	97
47) Tetrachloroethene	8.22	164	34619	5.831 ug/L	99
65) 1,2-Dichlorobenzene	11.88	146	2736	0.214 ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU070219\
Data File : VU033144.D
Acq On : 02 Jul 2019 16:34
Operator : JC/SP
Sample : K3608-07
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GAML6

Quant Time: Jul 03 03:00:53 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR070219WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Jul 03 02:14:33 2019
Response via : Initial Calibration

