

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU070219\
 Data File : VU033150.D
 Acq On : 02 Jul 2019 18:59
 Operator : JC/SP
 Sample : K3597-15DL 5X
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GAM72DL

Quant Time: Jul 03 03:57:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR070219WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Jul 03 02:14:33 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	86579	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.08	117	90045	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	41190	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	34115	6.81	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	136.20%#
7) Chloroethane-d5	1.68	69	28253	6.16	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	123.20%
11) 1,1-Dichloroethene-d2	2.27	63	53293	4.51	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	90.20%
20) 2-Butanone-d5	4.19	46	90891	49.82	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	99.64%
24) Chloroform-d	4.64	84	64559	5.49	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	109.80%
26) 1,2-Dichloroethane-d4	5.30	65	37112	5.47	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	109.40%
32) Benzene-d6	5.33	84	119264	5.52	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	110.40%
36) 1,2-Dichloropropane-d6	6.32	67	39469	5.49	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	109.80%
41) Toluene-d8	7.56	98	103246	4.97	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.40%
43) trans-1,3-Dichloropropene-	7.85	79	15217	4.92	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	98.40%
46) 2-Hexanone-d5	8.31	63	59292	46.52	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.04%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	31583	5.00	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.00%
64) 1,2-Dichlorobenzene-d4	11.85	152	41712	5.63	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	112.60%

Target Compounds					Ovalue
13) Acetone	2.32	43	3366	2.295 ug/L	90
47) Tetrachloroethene	8.22	164	54154	9.006 ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU070219\
Data File : VU033150.D
Acq On : 02 Jul 2019 18:59
Operator : JC/SP
Sample : K3597-15DL 5X
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GAM72DL

Quant Time: Jul 03 03:57:20 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR070219WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Jul 03 02:14:33 2019
Response via : Initial Calibration

