

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070221\
 Data File : VU044300.D
 Acq On : 02 Jul 2021 13:44
 Operator : SY/MD
 Sample : M2914-12
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DBK41

Quant Time: Jul 03 04:59:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR062121WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Jul 03 04:56:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.256	114	130656	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.423	117	123134	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.819	152	66134	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	27116	3.833	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	76.600%
7) Chloroethane-d5	1.919	69	30783	4.419	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	88.400%
11) 1,1-Dichloroethene-d2	2.575	65	14865	4.094	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	81.800%
20) 2-Butanone-d5	4.646	46	138154	53.035	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	106.060%
24) Chloroform-d	5.073	84	75898	4.766	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.400%
26) 1,2-Dichloroethane-d4	5.710	65	45337	4.763	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.200%
32) Benzene-d6	5.736	84	148979	4.871	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.400%
36) 1,2-Dichloropropane-d6	6.700	67	49552	5.000	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.000%
41) Toluene-d8	7.906	98	139696	4.923	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.400%
43) trans-1,3-Dichloroprop...	8.186	79	20118	4.376	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.600%
46) 2-Hexanone-d5	8.642	63	115646	52.817	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	105.640%
56) 1,1,2,2-Tetrachloroeth...	10.764	84	43447	4.931	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.600%
66) 1,2-Dichlorobenzene-d4	12.198	152	58045	5.327	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	106.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070221\
Data File : VU044300.D
Acq On : 02 Jul 2021 13:44
Operator : SY/MD
Sample : M2914-12
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DBK41

Quant Time: Jul 03 04:59:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR062121WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jul 03 04:56:00 2021
Response via : Initial Calibration

