

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070221\
 Data File : VU044302.D
 Acq On : 02 Jul 2021 14:32
 Operator : SY/MD
 Sample : VIBLK133
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK001

Quant Time: Jul 03 05:00:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR062121WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Jul 03 04:56:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.257	114	135102	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.424	117	128512	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.819	152	67713	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	30493	4.169	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	83.400%
7) Chloroethane-d5	1.919	69	33925	4.710	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.200%
11) 1,1-Dichloroethene-d2	2.575	65	16471	4.387	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	87.800%
20) 2-Butanone-d5	4.643	46	110608	41.063	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	82.120%
24) Chloroform-d	5.073	84	77052	4.679	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.600%
26) 1,2-Dichloroethane-d4	5.710	65	46816	4.756	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.200%
32) Benzene-d6	5.736	84	158196	4.956	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.200%
36) 1,2-Dichloropropane-d6	6.697	67	51563	4.985	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.600%
41) Toluene-d8	7.906	98	148459	5.013	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.200%
43) trans-1,3-Dichloroprop...	8.186	79	19291	4.021	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.400%
46) 2-Hexanone-d5	8.646	63	119512	52.298	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	104.600%
56) 1,1,2,2-Tetrachloroeth...	10.765	84	45225	4.918	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.400%
66) 1,2-Dichlorobenzene-d4	12.199	152	59609	5.343	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	106.800%
Target Compounds						Qvalue
13) Acetone	2.649	43	2775	2.304	ug/L	87
16) Methylene chloride	3.054	84	14283	1.358	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070221\
Data File : VU044302.D
Acq On : 02 Jul 2021 14:32
Operator : SY/MD
Sample : VIBLK133
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK001

Quant Time: Jul 03 05:00:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR062121WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jul 03 04:56:00 2021
Response via : Initial Calibration

