

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU070318\
 Data File : VU025190.D
 Acq On : 03 Jul 2018 14:52
 Operator : MD/SY
 Sample : J3760-01
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 WP021MW011-180627

Quant Time: Jul 05 09:01:30 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	203596	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	305964	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	280485	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	144580	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	155581	46.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.62%	
35) Dibromofluoromethane	4.89	113	119699	47.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.28%	
50) Toluene-d8	7.57	98	434985	46.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.92%	
62) 4-Bromofluorobenzene	10.31	95	159934	43.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.98%	
Target Compounds						
4) Vinyl Chloride	1.40	62	12405	4.620	ug/l	98
27) cis-1,2-Dichloroethene	4.24	96	18147	6.182	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU070318\
Data File : VU025190.D
Acq On : 03 Jul 2018 14:52
Operator : MD/SY
Sample : J3760-01
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WP021MW011-180627

Quant Time: Jul 05 09:01:30 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

