

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU070318\
 Data File : VU025191.D
 Acq On : 03 Jul 2018 15:17
 Operator : MD/SY
 Sample : J3760-04
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DUP-0450-180627

Quant Time: Jul 05 09:01:57 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	201316	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	302099	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	276106	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	137263	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	152727	46.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.94%	
35) Dibromofluoromethane	4.89	113	117871	47.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.02%	
50) Toluene-d8	7.57	98	429538	46.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.94%	
62) 4-Bromofluorobenzene	10.31	95	159120	43.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.64%	
Target Compounds						
4) Vinyl Chloride	1.40	62	11922	4.491	ug/l	100
27) cis-1,2-Dichloroethene	4.23	96	18594	6.406	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU070318\
Data File : VU025191.D
Acq On : 03 Jul 2018 15:17
Operator : MD/SY
Sample : J3760-04
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DUP-0450-180627

Quant Time: Jul 05 09:01:57 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

