

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU070318\
 Data File : VU025193.D
 Acq On : 03 Jul 2018 16:06
 Operator : MD/SY
 Sample : J3783-03
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 S10-TB-2018-2

Quant Time: Jul 04 05:41:36 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	203120	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	302970	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	280585	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	141412	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	153374	46.25	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.50%	
35) Dibromofluoromethane	4.89	113	118373	47.08	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.16%	
50) Toluene-d8	7.57	98	433193	47.23	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.46%	
62) 4-Bromofluorobenzene	10.31	95	157926	43.37	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.74%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU070318\
Data File : VU025193.D
Acq On : 03 Jul 2018 16:06
Operator : MD/SY
Sample : J3783-03
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
S10-TB-2018-2

Quant Time: Jul 04 05:41:36 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

