

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU070318\
 Data File : VU025199.D
 Acq On : 03 Jul 2018 19:15
 Operator : MD/SY
 Sample : J3840-04
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TP-6-S

Quant Time: Jul 04 06:32:18 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	200092	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	303117	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	278852	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	147526	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.31	65	153222	46.91	ug/l	0.00
Spiked Amount			Recovery	=	93.82%	
35) Dibromofluoromethane	4.89	113	117879	46.86	ug/l	0.00
Spiked Amount			Recovery	=	93.72%	
50) Toluene-d8	7.57	98	433791	47.27	ug/l	0.00
Spiked Amount			Recovery	=	94.54%	
62) 4-Bromofluorobenzene	10.31	95	161117	44.22	ug/l	0.00
Spiked Amount			Recovery	=	88.44%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.32	43	10384	5.916	ug/l	96
18) Methyl Acetate	2.62	43	19529	5.190	ug/l	99
20) Methylene Chloride	2.71	84	138126	50.157	ug/l	97
25) 2-Butanone	4.27	43	6641	2.817	ug/l	95
78) n-propylbenzene	10.60	91	2763	0.207	ug/l	95
84) 1,2,4-Trimethylbenzene	11.15	105	6936	0.693	ug/l	96
95) Naphthalene	13.74	128	328820	30.580	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU070318\
Data File : VU025199.D
Acq On : 03 Jul 2018 19:15
Operator : MD/SY
Sample : J3840-04
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TP-6-S

Quant Time: Jul 04 06:32:18 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

